

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1041
CALIBRATION DATE: 19-Sep-07

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.10144610e+000
h = 5.60571897e-001
i = 1.26571908e-004
j = 2.60637448e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 5.58765285e-005
b = 5.60910237e-001
c = -4.10266586e+000
d = -9.04815944e-005
m = 3.8
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70363	0.00000	0.00000
-1.0002	34.9324	2.81304	7.56622	2.81303	-0.00001
0.9998	34.9324	2.98492	7.76467	2.98494	0.00001
14.9998	34.9342	4.28452	9.12454	4.28450	-0.00003
18.4998	34.9339	4.63226	9.45483	4.63227	0.00001
28.9998	34.9321	5.71912	10.41856	5.71915	0.00002
32.4998	34.9261	6.09294	10.72959	6.09293	-0.00002

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

