

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1041  
CALIBRATION DATE: 10-Oct-07

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.10230250e+000  
h = 5.60832253e-001  
i = 6.50551589e-005  
j = 2.85408161e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 5.45981960e-005  
b = 5.60867856e-001  
c = -4.10200263e+000  
d = -7.79737532e-005  
m = 3.8  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70364	0.00000	0.00000
-1.0002	34.9689	2.81570	7.56988	2.81571	0.00001
0.9998	34.9695	2.98779	7.76848	2.98779	-0.00000
14.9998	34.9707	4.28852	9.12927	4.28851	-0.00001
18.4998	34.9710	4.63665	9.45981	4.63663	-0.00002
28.9998	34.9688	5.72445	10.42420	5.72452	0.00007
32.4998	34.9616	6.09843	10.73522	6.09839	-0.00004

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

