

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1054
CALIBRATION DATE: 11-Oct-07

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.08886663e+000
h = 5.91324562e-001
i = 2.44854437e-006
j = 3.54984170e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.49588219e-005
b = 5.91382889e-001
c = -4.08936130e+000
d = -8.94191972e-005
m = 4.0
CPcor = -9.5700e-008 (nominal)

| BATH TEMP (ITS-90) | BATH SAL (PSU) | BATH COND (Siemens/m) | INST FREQ (kHz) | INST COND (Siemens/m) | RESIDUAL (Siemens/m) |
|-----------------------|-------------------|--------------------------|--------------------|--------------------------|-------------------------|
| 0.0000 | 0.0000 | 0.00000 | 2.62903 | 0.00000 | 0.00000 |
| -1.0002 | 34.9575 | 2.81487 | 7.37147 | 2.81487 | 0.00000 |
| 0.9998 | 34.9576 | 2.98687 | 7.56493 | 2.98688 | 0.00001 |
| 14.9998 | 34.9585 | 4.28719 | 8.89050 | 4.28714 | -0.00004 |
| 18.4998 | 34.9581 | 4.63512 | 9.21246 | 4.63514 | 0.00001 |
| 28.9998 | 34.9542 | 5.72233 | 10.15157 | 5.72239 | 0.00006 |
| 32.4998 | 34.9454 | 6.09593 | 10.45431 | 6.09589 | -0.00004 |

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

