

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1070
CALIBRATION DATE: 18-Jul-13

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.10559895e+000
h = 5.93735069e-001
i = -7.40863298e-005
j = 3.60265781e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 2.39335465e-005
b = 5.93554238e-001
c = -4.10527984e+000
d = -8.55268833e-005
m = 4.1
CPcor = -9.5700e-008 (nominal)

| BATH TEMP (ITS-90) | BATH SAL (PSU) | BATH COND (Siemens/m) | INST FREQ (kHz) | INST COND (Siemens/m) | RESIDUAL (Siemens/m) |
|-----------------------|-------------------|--------------------------|--------------------|--------------------------|-------------------------|
| 0.0000 | 0.0000 | 0.00000 | 2.62949 | 0.00000 | 0.00000 |
| -1.0000 | 34.8468 | 2.80680 | 7.35260 | 2.80681 | 0.00001 |
| 1.0001 | 34.8476 | 2.97839 | 7.54557 | 2.97839 | 0.00000 |
| 15.0001 | 34.8492 | 4.27523 | 8.86768 | 4.27523 | -0.00001 |
| 18.5000 | 34.8497 | 4.62232 | 9.18884 | 4.62229 | -0.00004 |
| 29.0001 | 34.8487 | 5.70704 | 10.12610 | 5.70713 | 0.00009 |
| 32.5001 | 34.8447 | 6.08039 | 10.42870 | 6.08034 | -0.00006 |

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

