

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1538
CALIBRATION DATE: 06-Jul-06

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.20720840e+000
h = 4.70658801e-001
i = -1.10741140e-004
j = 2.68332205e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.20374729e-005
b = 4.70305796e-001
c = -4.20600200e+000
d = -8.24923572e-005
m = 4.2
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.99010	0.00000	0.00000
-0.9948	34.8804	2.80970	8.27659	2.80971	0.00001
1.0763	34.8806	2.98756	8.50071	2.98756	-0.00000
14.9999	34.8815	4.27875	9.97604	4.27871	-0.00004
18.4999	34.8808	4.62599	10.33646	4.62603	0.00004
28.9999	34.8793	5.71146	11.38813	5.71146	0.00000
32.4999	34.8719	6.08458	11.72738	6.08457	-0.00000


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 28-Oct-04 1.0000915
▲ 06-Jul-06 1.0000000