

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1538
CALIBRATION DATE: 17-Jun-09

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.21468637e+000
h = 4.71685562e-001
i = -1.95273222e-004
j = 3.00002543e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 5.04256345e-006
b = 4.71112804e-001
c = -4.21324029e+000
d = -9.06789202e-005
m = 4.5
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.99021	0.00000	0.00000
-1.0000	34.6282	2.79083	8.24869	2.79086	0.00003
1.0000	34.6290	2.96148	8.46430	2.96144	-0.00003
15.0000	34.6301	4.25118	9.94243	4.25115	-0.00003
18.5000	34.6296	4.59627	10.30154	4.59631	0.00004
29.0000	34.6290	5.67508	11.34951	5.67507	-0.00001
32.5000	34.6244	6.04630	11.68794	6.04630	0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

