

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1553  
CALIBRATION DATE: 03-Aug-12

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.18047385e+000  
h = 5.25270384e-001  
i = 5.82174689e-005  
j = 2.47461442e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 3.59095430e-005  
b = 5.25467807e-001  
c = -4.18141083e+000  
d = -8.93697342e-005  
m = 3.9  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.82014	0.00000	0.00000
-0.9999	34.7214	2.79765	7.80979	2.79771	0.00006
1.0000	34.7225	2.96871	8.01402	2.96867	-0.00004
15.0001	34.7240	4.26150	9.41427	4.26143	-0.00006
18.5000	34.7240	4.60744	9.75451	4.60746	0.00001
29.0001	34.7234	5.68882	10.74745	5.68894	0.00012
32.5001	34.7178	6.06076	11.06779	6.06068	-0.00008

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

