

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1568  
CALIBRATION DATE: 15-Feb-07

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.00318578e+000  
h = 5.18615499e-001  
i = -8.86761830e-005  
j = 2.86026889e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 1.79095061e-005  
b = 5.18308133e-001  
c = -4.00204279e+000  
d = -7.92553815e-005  
m = 4.1  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.77838	0.00000	0.00000
-1.0001	34.8567	2.80752	7.85662	2.80752	-0.00000
1.0160	34.8573	2.98052	8.06513	2.98053	0.00000
14.9999	34.8571	4.27608	9.48023	4.27606	-0.00001
18.4999	34.8560	4.62306	9.82422	4.62308	0.00002
28.9999	34.8532	5.70767	10.82792	5.70765	-0.00002
32.4999	34.8451	6.08043	11.15165	6.08044	0.00001

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

