

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1568  
CALIBRATION DATE: 25-Feb-14

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.00234279e+000  
h = 5.18426941e-001  
i = -4.77847737e-005  
j = 2.69354274e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 1.82013209e-005  
b = 5.18336141e-001  
c = -4.00239205e+000  
d = -8.89146817e-005  
m = 4.1  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.77832	0.00000	0.00000
-1.0000	34.8083	2.80399	7.85199	2.80399	0.00000
1.0000	34.8087	2.97538	8.05865	2.97537	-0.00000
15.0000	34.8083	4.27074	9.47433	4.27075	0.00001
18.5000	34.8080	4.61739	9.81810	4.61738	-0.00001
29.0000	34.8065	5.70089	10.82127	5.70088	-0.00001
32.5000	34.7999	6.07345	11.14497	6.07346	0.00001


Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 28-Nov-12 1.0001439  
▲ 25-Feb-14 1.0000000