

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1896
CALIBRATION DATE: 15-Feb-07

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.10120991e+000
h = 5.23977666e-001
i = -1.15320660e-003
j = 8.67540338e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.47946389e-007
b = 5.19448194e-001
c = -4.08430192e+000
d = -5.90741620e-005
m = 6.0
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.80453	0.00000	0.00000
-1.0001	34.8567	2.80752	7.86412	2.80749	-0.00003
1.0160	34.8573	2.98052	8.07250	2.98056	0.00004
14.9999	34.8571	4.27608	9.48560	4.27606	-0.00002
18.4999	34.8560	4.62306	9.82890	4.62307	0.00001
28.9999	34.8532	5.70767	10.82982	5.70768	0.00001
32.4999	34.8451	6.08043	11.15233	6.08043	-0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

