

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1898  
CALIBRATION DATE: 06-Jul-06

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -1.05855343e+001  
h = 1.51684832e+000  
i = 3.55424706e-004  
j = 8.31147808e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 2.78891996e-004  
b = 1.51703975e+000  
c = -1.05854534e+001  
d = -8.26160174e-005  
m = 3.6  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.64039	0.00000	0.00000
-0.9948	34.8804	2.80970	5.04345	2.80971	0.00001
1.0763	34.8806	2.98756	5.15786	2.98755	-0.00001
14.9999	34.8815	4.27875	5.92216	4.27872	-0.00003
18.4999	34.8808	4.62599	6.11132	4.62603	0.00004
28.9999	34.8793	5.71146	6.66767	5.71145	-0.00001
32.4999	34.8719	6.08458	6.84839	6.08458	0.00000

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

