

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1898  
CALIBRATION DATE: 11-Aug-09

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -1.05822116e+001  
h = 1.51657794e+000  
i = 3.45478423e-004  
j = 5.11192168e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 3.20319842e-004  
b = 1.51650899e+000  
c = -1.05815285e+001  
d = -8.12622857e-005  
m = 3.4  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.64043	0.00000	0.00000
-1.0000	34.7667	2.80095	5.03941	2.80095	-0.00000
1.0000	34.7671	2.97216	5.14984	2.97218	0.00002
15.0000	34.7687	4.26639	5.91795	4.26638	-0.00001
18.5000	34.7689	4.61276	6.10706	4.61274	-0.00002
29.0000	34.7669	5.69514	6.66336	5.69521	0.00007
32.5000	34.7623	6.06764	6.84422	6.06759	-0.00005

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

