

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA  
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1898  
 CALIBRATION DATE: 25-Feb-14

SBE4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

### GHIJ COEFFICIENTS

g = -1.05861318e+001  
 h = 1.51744281e+000  
 i = 1.62331862e-004  
 j = 6.93574688e-005  
 CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

### ABCDM COEFFICIENTS

a = 1.59123806e-004  
 b = 1.51754238e+000  
 c = -1.05862665e+001  
 d = -8.47728199e-005  
 m = 3.7  
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.64047	0.00000	0.00000
-1.0000	34.8083	2.80399	5.04097	2.80399	-0.00000
1.0000	34.8087	2.97538	5.15144	2.97537	-0.00001
15.0000	34.8083	4.27074	5.91978	4.27075	0.00002
18.5000	34.8080	4.61739	6.10891	4.61738	-0.00000
29.0000	34.8065	5.70089	6.66535	5.70086	-0.00003
32.5000	34.7999	6.07345	6.84618	6.07347	0.00002

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

