

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2356
CALIBRATION DATE: 06-Jul-06

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.00798475e+001
h = 1.46971711e+000
i = 4.45233043e-004
j = 6.17583037e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.10995579e-004
b = 1.47002302e+000
c = -1.00800823e+001
d = -8.38391168e-005
m = 3.5
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.61743	0.00000	0.00000
-0.9948	34.8804	2.80970	5.08994	2.80971	0.00001
1.0763	34.8806	2.98756	5.20693	2.98755	-0.00001
14.9999	34.8815	4.27875	5.98777	4.27874	-0.00002
18.4999	34.8808	4.62599	6.18086	4.62603	0.00004
28.9999	34.8793	5.71146	6.74851	5.71142	-0.00005
32.4999	34.8719	6.08458	6.93286	6.08460	0.00003

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

