

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2356  
CALIBRATION DATE: 19-Jul-11

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -1.00758804e+001  
h = 1.46758423e+000  
i = 1.11929059e-003  
j = 2.13760395e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 1.34630946e-003  
b = 1.46692566e+000  
c = -1.00743673e+001  
d = -8.07672812e-005  
m = 3.0  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.61749	0.00000	0.00000
-1.0000	34.8725	2.80868	5.08854	2.80866	-0.00002
1.0000	34.8731	2.98036	5.20152	2.98039	0.00004
15.0000	34.8740	4.27794	5.98616	4.27792	-0.00002
18.5000	34.8739	4.62518	6.17915	4.62516	-0.00002
29.0000	34.8710	5.71027	6.74646	5.71034	0.00007
32.5000	34.8637	6.08332	6.93061	6.08327	-0.00005

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

