

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0387
 CALIBRATION DATE: 16-Nov-06p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.4613

Boc = 0.0000

Voffset = -0.4881

TCor = 0.0010

PCor = 1.350e-04

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.23	2.00	0.00	0.765	1.24	0.01
1.23	6.00	0.00	0.795	1.24	0.01
1.25	12.00	0.01	0.844	1.25	0.01
1.27	20.00	0.01	0.910	1.26	-0.00
1.28	26.00	0.01	0.962	1.27	-0.01
1.30	30.00	0.01	1.002	1.29	-0.01
4.10	12.00	0.01	1.657	4.11	0.02
4.11	30.00	0.01	2.123	4.10	-0.01
4.11	6.00	0.00	1.506	4.11	0.00
4.11	26.00	0.01	2.016	4.10	-0.01
4.12	20.00	0.01	1.867	4.12	0.00
4.13	2.00	0.00	1.409	4.11	-0.02
6.72	12.00	0.01	2.406	6.75	0.02
6.74	6.00	0.00	2.159	6.75	0.01
6.76	2.00	0.00	1.998	6.74	-0.02
6.78	20.00	0.01	2.755	6.78	0.00
6.81	30.00	0.01	3.206	6.82	0.01
6.81	26.00	0.01	3.021	6.80	-0.02

$$\text{oxygen (ml/l)} = (\text{Soc} * (\text{V} + \text{Voffset})) * \exp(\text{Tcor} * \text{T}) * \text{Oxsat}(\text{T},\text{S}) * \exp(\text{Pcor} * \text{P})$$

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU]

Oxsat(T,S) = oxygen saturation [ml/l], P = pressure [dbar]

Residual = instrument oxygen - bath oxygen

