

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0387
CALIBRATION DATE: 30-Jul-08p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.3932

Voffset = -0.4835

Tau20 = 1.12

A = 5.6088e-004

B = 4.8526e-005

C = -1.5281e-006

E nominal = 0.036

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.24	2.00	0.00	0.809	1.24	0.00
1.25	6.00	0.00	0.848	1.25	0.00
1.26	12.00	0.01	0.903	1.26	0.00
1.26	20.00	0.01	0.978	1.26	0.00
1.26	26.00	0.01	1.038	1.26	-0.00
1.27	30.00	0.01	1.083	1.27	0.00
4.13	30.00	0.01	2.432	4.13	-0.00
4.14	26.00	0.01	2.300	4.14	-0.00
4.16	20.00	0.01	2.115	4.16	0.00
4.18	2.00	0.00	1.579	4.17	-0.00
4.18	6.00	0.00	1.700	4.19	0.00
4.19	12.00	0.01	1.880	4.19	-0.00
6.51	30.00	0.01	3.557	6.51	0.00
6.59	26.00	0.01	3.379	6.60	0.01
6.66	20.00	0.01	3.099	6.66	-0.00
6.70	12.00	0.01	2.719	6.70	-0.00
6.72	2.00	0.00	2.248	6.72	-0.00
6.75	6.00	0.00	2.446	6.75	0.01

$$\text{Oxygen (ml/l)} = \text{Soc} * (\text{V} + \text{Voffset}) * (1.0 + \text{A} * \text{T} + \text{B} * \text{T}^2 + \text{C} * \text{T}^3) * \text{OxSol}(\text{T},\text{S}) * \exp(\text{E} * \text{P} / \text{K})$$

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)

