



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 68054

Date 4/24/2012

Technician JK

Serial Number 4556968-0360

Low Pressure (PSI) 30 PSI

Time (Minutes) 30 Minutes

High Pressure (PSI) N/A * PSI

Time (Minutes) N/A * Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.
*The unit is not designed for high pressure applications.

