Temperature Calibration Report

Customer:	Oregon State Ur	niversity			
Job Number:	68054	Date of Re	port:	4/23/2012	
Model Number:	SBE 45	Serial Nun	nber:	4556968-0360	
Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request. An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.					
'AS RECEIVED CALIBRATION'		<u>✓</u> P	✓ Performed □ Not Performed		
Date: 3/14/2012]	Drift since last cal:	-0.0	Degrees Celsius/year	
Comments:					
'CALIBRATION A	AFTER REPAIR'	□ P	erformed	d ✓ Not Performed	
Date:]	Drift since Last cal:		Degrees Celsius/year	
Comments:					