



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 65915

Date 9/28/2011

Technician JK

Serial Number 4556968-0361

Low Pressure (PSI) 30 PSI

Time (Minutes) 30 Minutes

High Pressure (PSI) PSI

Time (Minutes) Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. *The unit is not designed for high pressure applications.

