



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer Oregon State University

Job Number 68054

Date 4/3/2012

Technician BA

Serial Number 480047

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) N/A \* PSI

Time (Minutes) N/A \* Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.  
\*The unit is not designed for high pressure applications.

