



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 71811

Date 12/12/2012

Technician AL

Serial Number 480047

Low Pressure (PSI) 45 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 45 PSI

Time (Minutes) 15 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. *The unit is not designed for high pressure applications.

