

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0047  
CALIBRATION DATE: 22-Feb-14

SBE 48 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

a0 = -9.006649e-006  
a1 = 2.738624e-004  
a2 = -2.450997e-006  
a3 = 1.488791e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	837167.2	1.0000	0.0000
4.5000	714225.2	4.5000	-0.0000
15.0000	452343.0	15.0000	0.0000
18.5000	390900.7	18.5000	0.0000
24.0000	312636.4	24.0000	0.0000
28.9999	256740.9	28.9999	-0.0000
32.5000	224419.8	32.5000	0.0000

Temperature ITS-90 =  $1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$  (°C)

Residual = instrument temperature - bath temperature

