



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer Oregon State University

Job Number 63973

Date 6/15/2011

Technician AD

Serial Number 09P5745-0250

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 60 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.

