



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service	RMA Number	51445A
Report		

Customer Information:

Company	Oregon State University	Date	10/2/2008
Contact	Dave O Gorman		
PO Number	S0953A		
Serial Number	09P6012-0256		
Model Number	SBE 09Plus		

Services Requested:

1. Evaluate/Repair Instrumentation.

Problems Found:

1. 9P voltage channels were reading zero and frequency 0 was reading a signal without anything connected to JB1. Found U11 on AP counter card for Temperature 1 reading faulty. Replaced U11.

Services Performed:

1. Performed initial diagnostic evaluation.
2. Performed hydrostatic pressure test.
3. Calibrated the pressure sensor.
4. Performed complete system check and full diagnostic evaluation.

Special Notes:

--