



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 79433

Date 8/6/2014

Technician PS

Serial Number 09P13334-0428

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 45 Minutes

Pass

Fail

Comments

Lower endcap, connectors, and all orings replaced.

