



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

**Customer**      ***Oregon State University***

**Job Number**      **57417**

**Date**              2/9/2010

**Technician**      SF

<b><u>Serial Number</u></b>	92843
-----------------------------	-------

**Low Pressure (PSI)**      50      PSI

**Time (Minutes)**              15      Minutes

**High Pressure (PSI)**      10000      PSI

**Time (Minutes)**              30      Minutes

Pass     

Fail     

Comments

Replaced the main piston "O"-Rings. Replaced 6-pin connector.

