



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 78604

Date 7/17/2014

Technician PS

Serial Number 92843

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 9000 PSI

Time (Minutes) 45 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. Unit converted to wet plugable connectors.

