



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service
Report

RMA Number	49859
-------------------	-------

Customer Information:

Company	Oregon State University	Date	3/31/2008
----------------	-------------------------	-------------	-----------

Contact	Dave O Gorman
----------------	---------------

PO Number	S0953A
------------------	--------

Serial Number	430463
Model Number	SBE 43

Services Requested:

1. Perform Routine Calibration Service.

Problems Found:

--

Services Performed:

1. Performed initial diagnostic evaluation. 2. Performed "Post Cruise" calibration of the oxygen sensor. 3. Performed full diagnostic evaluation.

Special Notes:

--