



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service
Report

RMA Number

65218

Customer Information:

Company	Oregon State University	Date	9/21/2011
Contact	Dave O Gorman		
PO Number	S1180A		

Serial Number	92843
Model Number	SBE 09Plus

Services Requested:

1. Evaluate/Repair Instrumentation.
2. Perform Routine Calibration Service.

Problems Found:

1. The power supply and other components on the Sea Cable interface were found to have failed.

Services Performed:

1. Performed initial diagnostic evaluation.
2. Installed NEW power supply.
3. Performed internal inspection and O-ring replacement.
4. Performed hydrostatic pressure test.
5. Calibrated the pressure sensor.
6. Performed complete system check and full diagnostic evaluation.

Special Notes:

--