



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0080  
 CALIBRATION DATE: 26-Feb-19

SBE 45 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.897703e-001      CPcor = -9.5700e-008  
 h = 1.434755e-001      CTcor = 3.2500e-006  
 i = -2.384454e-004      WBOTC = 9.3817e-006  
 j = 4.248022e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2629.29	0.00000	0.00000
1.0000	34.6623	2.96405	5250.99	2.96406	0.00001
4.5000	34.6429	3.26997	5449.45	3.26996	-0.00001
15.0000	34.6020	4.24810	6039.49	4.24809	-0.00001
18.5000	34.5934	4.59198	6233.34	4.59198	0.00000
23.9999	34.5838	5.14784	6534.16	5.14785	0.00002
29.0000	34.5784	5.66772	6803.06	5.66771	-0.00001
32.5000	34.5742	6.03853	6988.27	6.03845	-0.00008

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars);  $\delta = \text{CTcor}$ ;  $\epsilon = \text{CPcor}$ ;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

