



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3364
 CALIBRATION DATE: 20-Mar-20

SBE 21 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -4.03406162e+000 CPcor = -9.5700e-008 (nominal)
 h = 4.74803994e-001 CTcor = 3.2500e-006 (nominal)
 i = 2.85534746e-005
 j = 2.38363126e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.91396	0.00000	0.00000
1.0000	34.7340	2.96960	8.41140	2.96953	-0.00007
4.5000	34.7146	3.27607	8.78397	3.27617	0.00010
15.0000	34.6737	4.25597	9.87926	4.25592	-0.00005
18.5000	34.6652	4.60048	10.23610	4.60044	-0.00004
24.0000	34.6555	5.15734	10.78760	5.15746	0.00011
29.0000	34.6483	5.67789	11.27790	5.67780	-0.00009
32.5000	34.6414	6.04893	11.61469	6.04896	0.00003

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

