

BGM3 zero-damp test report

HERR, vessel: SALLY RIDE

Release Date: 2017/01/25 18:24:52 UTC

Sensor: S224

Software version: 1.2

Port/Pier/Berth: MARFAC

Axis	pitch
Periods	4
Test start	2017/01/25 18:15:39.298
T1	2017/01/25 18:17:52.949
T2	2017/01/25 18:20:07.309
T3	2017/01/25 18:22:21.065
Test end	2017/01/25 18:24:35.364
Tolerance	5 s
4T Error	3.066 s

Table 1: Timings

Action: Performance within specification and tolerance. No action required.