

TABLE 1

MILLIGAL VALUES FOR LACOSTE & ROMBERG, INC. MODEL G GRAVITY METER #G- 611

COUNTER READING*	VALUE IN MILLIGALS	FACTOR FOR INTERVAL	COUNTER READING*	VALUE IN MILLIGALS	FACTOR FOR INTERVAL
000	000.00	1.01998			
100	102.00	1.01994	3600	3673.37	1.02170
200	203.99	1.01990	3700	3775.54	1.02175
300	305.98	1.01986	3800	3877.72	1.02181
400	407.97	1.01982	3900	3979.90	1.02187
500	509.95	1.01977	4000	4082.09	1.02195
600	611.93	1.01975	4100	4184.28	1.02201
700	713.90	1.01973	4200	4286.48	1.02206
800	815.88	1.01973	4300	4388.69	1.02212
900	917.85	1.01976	4400	4490.90	1.02216
1000	1019.82	1.01977	4500	4593.12	1.02221
1100	1121.80	1.01980	4600	4695.34	1.02223
1200	1223.78	1.01983	4700	4797.56	1.02226
1300	1325.76	1.01988	4800	4899.79	1.02228
1400	1427.75	1.01994	4900	5002.01	1.02230
1500	1529.75	1.02001	5000	5104.24	1.02230
1600	1631.75	1.02007	5100	5206.47	1.02231
1700	1733.75	1.02012	5200	5308.70	1.02230
1800	1835.77	1.02021	5300	5410.93	1.02226
1900	1937.79	1.02027	5400	5513.16	1.02223
2000	2039.81	1.02034	5500	5615.38	1.02219
2100	2141.85	1.02041	5600	5717.60	1.02211
2200	2243.89	1.02050	5700	5819.81	1.02204
2300	2345.94	1.02058	5800	5922.02	1.02195
2400	2448.00	1.02065	5900	6024.21	1.02195
2500	2550.06	1.02073	6000	6126.40	1.02175
2600	2652.14	1.02083	6100	6228.57	1.02164
2700	2754.22	1.02092	6200	6330.74	1.02152
2800	2856.31	1.02102	6300	6432.89	1.02139
2900	2958.41	1.02111	6400	6535.03	1.02121
3000	3060.52	1.02121	6500	6637.15	1.02106
3100	3162.64	1.02130	6600	6739.25	1.02084
3200	3264.77	1.02138	6700	6841.34	1.02062
3300	3366.91	1.02145	6800	6943.40	1.02036
3400	3469.06	1.02154	6900	7045.44	1.02007
3500	3571.21	1.02161	7000	7147.44	

* Note: Right-hand wheel on counter indicates approximately 0.1 milligal.

5-28-81

rp

Reading Line: 2.80

Temperature: 49.1°