

## **Pressure Test Certificate**

Test Date: 2023-08-28 Description: SBE-9plus CTD

## **Sensor Information:**

Model Number: SBE-9P

Serial Number: 1549

## **Pressure Test Protocol:**

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 10000 PSI Held For: 15 Minutes

Passed Test: True

Tested By: AM

High pressure is generally equal to the maximum depth rating of the instrument

Pressure

**Typical Test Profile**