



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0289
 CALIBRATION DATE: 07-Feb-24

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.943507e-001 CPcor = -9.5700e-008
 h = 1.623842e-001 CTcor = 3.2500e-006
 i = -3.837880e-004 WBOTC = 3.3617e-007
 j = 5.632767e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2479.18	0.00000	0.00000
1.0000	34.6514	2.96321	4944.70	2.96321	0.00000
4.5000	34.6319	3.26904	5131.67	3.26903	-0.00000
15.0000	34.5900	4.24678	5687.62	4.24677	-0.00001
18.5000	34.5809	4.59050	5870.32	4.59053	0.00003
24.0000	34.5703	5.14606	6153.81	5.14604	-0.00002
29.0000	34.5630	5.66548	6407.25	5.66549	0.00001
32.5000	34.5561	6.03573	6581.69	6.03563	-0.00010

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

