



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0289
 CALIBRATION DATE: 11-Feb-24

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.027734e+000 CPcor = -9.5700e-008
 h = 1.678929e-001 CTcor = 3.2500e-006
 i = -4.050326e-004 WBOTC = 3.3617e-007
 j = 6.057448e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2478.80	0.00000	0.00000
1.0000	34.6324	2.96174	4882.41	2.96175	0.00001
4.5000	34.6128	3.26741	5065.46	3.26740	-0.00001
15.0000	34.5708	4.24467	5610.07	4.24465	-0.00002
18.5000	34.5616	4.58821	5789.11	4.58820	-0.00001
24.0000	34.5508	5.14348	6067.05	5.14353	0.00005
29.0000	34.5429	5.66256	6315.45	5.66253	-0.00002
32.5000	34.5335	6.03223	6486.26	6.03191	-0.00032

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

